

LIFETIME LIMITED WARRANTY

If you have a question about your MOBB Healthcare device or this warranty, please contact an authorized MOBB Healthcare dealer.

MOBB Healthcare warrants to the original consumer purchaser for as long as they own their MOBB product that it will be free from defects in material and workmanship under normal use. This warranty is only applicable to products used in residential applications within the United States and Canada and is not transferable to subsequent owners. Should defects in material or workmanship exist during the applicable warranty period in any MOBB Healthcare Product, MOBB Healthcare will elect at its discretion to either repair or replace the defective material. This warranty covers only parts and materials of the products supplied by MOBB Healthcare. Not covered under this warranty are wear and tear items or abuse of the product in a manner that was not intended for the product to be used. This warranty shall become void if the product is in any way modified, improperly adjusted or abused.

MHWTCD18



844604104029

MOBB Health Care Ltd.

116A Viceroy Rd.
Concord, ON L4K 2M1 Canada
Toll: 1-866-238-7190
Tel: 905-669-0348
Fax: 905-669-2796
www.mobbhhc.com



Aluminum lightweight wheelchair/transport chair Duo



INSTRUCTIONS INSIDE

Wheelchair/Transport chair Set up & Usage

1. Unfolding

- Locate the two rails on the seat rest. This is where the seat fabric is attached to. Push down on the seat rails and the chair will widen.

2. Fold

- Depress locking tabs on the back of the backrest "canes" to allow backrest to fold.

- Turn the right and left foot plates upward to the vertical position.

- Pull up on the middle section of the seat to fold.

3. Attach footrest

- Find the 2 pins on the wheelchair frame and place the 2 holes of the footrest on. Rotate the leg rest to align with the wheelchair frame and you should hear an audible click as it engages.

4. Adjustment of Foot Plate

- Loosen the screw of the foot rest shaft. Slide footplate up and down to desire height. Tighten after adjustment.

5. Transfer to Transport chair

- To transfer to transport chair, depress quick release push button on 24" wheel and slide out.



SAFETY FIRST

- Ensure the armrest is securely locked BEFORE using the wheelchair.
- Do not stand on the frame of the wheelchair.
- Always use the hand rims for self propulsion. Do not grip the rear wheels themselves.
- The wheelchair is suitable for indoors and out, but avoid prolonged bumpy terrain. Use proper caregiver techniques when negotiating obstacles.
- Consult with your MOBB Healthcare dealer for proper usage.
- The wheel locks are used while the occupant is sitting and stationary, or getting into or out of the wheel chair. The wheel locks are not meant to arrest the chair while moving. Doing so could cause injury and/or wheelchair damage.

